

**Safety Communication**

**رسالة سلامة**

**Software issues for the Infinity Acute Care System (IACS) and Standalone Infinity M540 patient monitor**

<b>Device/ Product Description:</b>	<ul style="list-style-type: none"> <li>The Infinity Acute Care System (IACS).</li> <li>Standalone Infinity M540 patient monitor.</li> </ul>
<b>Affected product:</b>	Software version VG7.1
<b>Manufacturer:</b>	Draeger Medical Systems
<b>Problem:</b>	<ol style="list-style-type: none"> <li>The Infinity M540 patient monitor may randomly reboot due to an error to correctly transmit and read the header data of files in the memory of the device.</li> <li>The IACS Cockpit may lose the ability to send commands to the M540 patient monitor.</li> <li>The IACS patient monitor may exhibit a loss of configuration of the patient and system profile due to disconnection from the internal database after power up of the Cockpit.</li> </ol>
<b>Recommendation /Actions:</b>	<p>Draeger will provide the affected customers with an updated software version VG 7.1.1 to resolve these product issues.</p> <p>For more information, please check “ <a href="#">NCMDR</a> “</p> <p>If you think you had a problem with your device or a device your patient uses, please report the problem to SFDA through:</p> <p><a href="#">NCMDR</a> <a href="#">Vigilance system</a> 19999 unified call center</p>

Devices/Products  
photo:



**Authorized  
Representative  
Details**

AR name:	Draeger Arabia Co. Ltd.
Assigned Contact Person:	Ahmed AlGarni
Mobile/Phone:	+966 583330000
Email:	Ahmed.Algarni@draeger.com