

Safety Alerts Weekly Update

التقرير الأسبوعي لإنذارات السلامة

Report Reference: WU2632
Publish date: 02-Aug-26

الرقم المرجعي للتقرير:
تاريخ النشر:

below is the weekly report of Safety Alerts for the period:

فيما يلي التقرير الأسبوعي لإنذارات السلامة للفترة:

From 26-Jul-26
To 01-Aug-26

من
إلى

which affect Saudi Arabia and being followed up with the authorised representatives to accomplish the required action.

والمتأثرة بها المملكة والتي جاري متابعتها مع الممثلين المعتمدين لإتمام تنفيذ الإجراءات التصحيحية.

* Kindly respond to the weekly report in both cases either you are affected or not affected though the following link:

* نأمل الرد على التقرير الأسبوعي في حالتي التأثر أو عدم التأثر وذلك من خلال الرابط أدناه:

<https://surveys.sfda.gov.sa/surveys/?s=7KN979YXNHPPRENX>



* Role of contact officer:

- Disseminate and share the information with other departments within the healthcare facility and ensure that the healthcare facility is free of any affected device/product.
- Communicate with the Authorised Representative of the manufacturer if there is any device/product affected by a Safety Alert
- To identify the affected serial numbers/lots, please open the Safety link.

* مسؤولية ضابط الاتصال:

- التعميم على الإدارات / الأقسام المختلفة داخل المنشأة الصحية والتأكد من خلوها من أي جهاز/مستلزم طبي متأثر بأي من إنذارات السلامة.
- التواصل مع الممثل المعتمد للمصنع في حالة وجود جهاز/مستلزم طبي متأثر بأي من إنذارات السلامة.
- لمعرفة تفاصيل الأجهزة والمستلزمات الطبية المتأثرة، الرجاء فتح رابط إنذار السلامة:



No. of Safety Alerts: 6 عدد إنذارات السلامة

Safety Alert No.	NCMDR Ref.	Medical Device	Manufacturer	Authorized Representative /Importer	Link	Medical Device Category
1	SA-22-07-26-1479	Panocell® -16	Immucor GmbH.	Medical supplies & Services Co.Ltd Mediserv	https://ade.sfda.gov.sa/Fsca/PublishD	In vitro diagnostic devices
2	SA-27-07-26-1482	L-Cath™ PICC S/L (Peripherally Inserted Central Catheter Basic Kit 28GA (1.2F) x 25cm 1 Lumen)	Argon Medical Devices, Inc.	Tart Medical	https://ade.sfda.gov.sa/Fsca/PublishD	Single-use devices
3	SA-26-07-26-1481	BECKMAN COULTER CHEMISTRY ANALYZER AU480	Beckman Coulter....	DHR Diagnostics Arabia Limited	https://ade.sfda.gov.sa/Fsca/PublishD	In vitro diagnostic devices
4	SA-29-07-26-1485	STA - Staclot Protein S	Diagnostica Stago S.A.S.	Diagnostic Saudi Arabia Trading Company	https://ade.sfda.gov.sa/Fsca/PublishD	In vitro diagnostic devices
5	SA-29-07-26-1484	STA® - C.K. Prest	Diagnostica Stago S.A.S.	Diagnostic Saudi Arabia Trading Company	https://ade.sfda.gov.sa/Fsca/PublishD	In vitro diagnostic devices
6	SA-30-07-26-1486	Insertion Kit, Guidewire, dilator.	Geringe Disinfection AB	Gulf Medical Co.	https://ade.sfda.gov.sa/Fsca/PublishD	Single-use devices